BenchPro, Inc.
Airborne Particle Count Record

Area: Test Chamber
Technician: Stefan Hocom
Testing Date: 2/15/2013

Test: Certification
Sample Volume: 1 ft^3

		Cumulative Counts/ft^3					Cumulative Counts/m^3				
Location #	**Time**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**
1	13:27:25	0	0	0	0	0	**0**	**0**	**0**	**0**	**0**
2	13:28:26	0	0	0	0	0	**0**	**0**	**0**	**0**	**0**
3	13:29:27	0	0	0	0	0	**0**	**0**	**0**	**0**	**0**

Remarks : Passed ISO Class 3 @ 0.1 micron

BenchPro, Inc.
Airborne Particle Count Record

Area: Test Chamber
Technician: Stefan Hocom
Testing Date: 2/15/2013

Test: Static Test 1
Sample Volume: 1 ft^3

		Cumulative Counts/ft^3					Cumulative Counts/m^3				
Location #	**Time**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**
1	13:33:30	17	2	0	0	0	**600**	**71**	**0**	**0**	**0**
2	13:34:40	10	5	1	0	0	**353**	**177**	**35**	**0**	**0**
3	13:37:00	13	4	2	1	1	**459**	**141**	**71**	**35**	**35**

Remarks : Passed ISO Class 3 @ 0.1 micron

BenchPro, Inc.
Airborne Particle Count Record

Area: Test Chamber
Technician: Stefan Hocom
Testing Date: 2/15/2013

Test: Walking Test
Sample Volume: 1 ft^3

		Cumulative Counts/ft^3					Cumulative Counts/m^3				
Location #	**Time**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**
1	13:39:00	123	100	90	73	66	**4342**	**3530**	**3177**	**2577**	**2330**
2	13:40:10	64	47	42	37	32	**2259**	**1659**	**1483**	**1306**	**1130**
3	13:41:20	97	73	64	57	51	**3424**	**2577**	**2259**	**2012**	**1800**

Remarks : Passed ISO Class 5 @ 0.5 micron

BenchPro, Inc.
Airborne Particle Count Record

Chair: LNT-VCR
Technician: Stefan Hocom
Testing Date: 2/15/2013

Test: Static 2
Sample Volume: 1 ft^3

		Cumulative Counts/ft^3					Cumulative Counts/m^3				
Location #	**Time**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**
1	14:51:47	0	0	0	0	0	**0**	**0**	**0**	**0**	**0**
2	14:53:02	0	0	0	0	0	**0**	**0**	**0**	**0**	**0**
3	14:54:17	0	0	0	0	0	**0**	**0**	**0**	**0**	**0**
4	14:55:32	0	0	0	0	0	**0**	**0**	**0**	**0**	**0**

Remarks : Indicates ISO class 3 compliance at 0.1 micron

BenchPro, Inc.
Airborne Particle Count Record

Chair: LNT-VCR
Technician: Stefan Hocom
Testing Date: 2/15/2013

Test: Static 3
Sample Volume: 1 ft^3

		Cumulative Counts/ft^3					Cumulative Counts/m^3				
Location #	**Time**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**
1	15:08:48	49	39	36	25	19	**1730**	**1377**	**1271**	**883**	**671**
2	15:10:03	38	29	26	18	15	**1341**	**1024**	**918**	**635**	**530**
3	15:11:18	26	17	15	7	5	**918**	**600**	**530**	**247**	**177**
4	15:12:33	14	8	6	3	2	**494**	**282**	**212**	**106**	**71**

Remarks : Indicates ISO class 3 compliance at 0.1 micron when control concentrations are removed (walking test).

BenchPro, Inc.
Airborne Particle Count Record

Chair: LNT-VCR
Technician: Stefan Hocom
Testing Date: 2/15/2013

Test: Engagement
Sample Volume: 1 ft^3

		Cumulative Counts/ft^3					Cumulative Counts/m^3				
Location #	**Time**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**
1	15:14:43	966	748	605	439	320	**34100**	**26404**	**21357**	**15497**	**11296**
2	15:15:58	923	571	436	303	210	**32582**	**20156**	**15391**	**10696**	**7413**
3	15:17:13	675	492	384	286	217	**23828**	**17368**	**13555**	**10096**	**7660**
4	15:18:28	1503	1171	959	681	489	**53056**	**41336**	**33853**	**24039**	**17262**

Remarks : Indicates ISO class 6 compliance at 0.5 micron.

BenchPro, Inc.
Airborne Particle Count Record

Chair: LNT-VCR
Technician: Stefan Hocom
Testing Date: 2/15/2013

Test: Mimic
Sample Volume: 1 ft^3

		Cumulative Counts/ft^3					Cumulative Counts/m^3				
Location #	**Time**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**
1	15:21:57	28	12	9	6	3	**988**	**424**	**318**	**212**	**106**
2	15:23:12	39	24	19	13	7	**1377**	**847**	**671**	**459**	**247**
3	15:24:27	46	24	15	12	6	**1624**	**847**	**530**	**424**	**212**
4	15:25:42	45	25	19	13	10	**1589**	**883**	**671**	**459**	**353**

Remarks : Indicates ISO class 3 compliance at 0.1 micron when control concentrations are removed (walking test).

BenchPro, Inc.
Airborne Particle Count Record

Chair: LNT-VCR
Technician: Stefan Hocom
Testing Date: 2/15/2013

Test: Adjustment
Sample Volume: 1 ft^3

		Cumulative Counts/ft^3					Cumulative Counts/m^3				
Location #	**Time**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**
1	15:27:16	1438	960	767	539	406	**50761**	**33888**	**27075**	**19027**	**14332**
2	15:28:31	534	389	317	230	161	**18850**	**13732**	**11190**	**8119**	**5683**
3	15:29:46	628	453	369	272	195	**22168**	**15991**	**13026**	**9602**	**6884**
4	15:31:01	730	304	218	149	119	**25769**	**10731**	**7695**	**5260**	**4201**

Remarks : Indicates ISO class 6 compliance at 0.5 micron.

BenchPro, Inc.
Airborne Particle Count Record

Chair: LNT-VCR
Technician: Stefan Hocom
Testing Date: 2/15/2013

Test: Recovery
Sample Volume: 1 ft^3

		Cumulative Counts/ft^3					Cumulative Counts/m^3				
Location #	**Time**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**	**0.1µm**	**0.2µm**	**0.3µm**	**0.5µm**	**1.0µm**
1	15:32:26	0	0	0	0	0	**0**	**0**	**0**	**0**	**0**
	15:32:32	1099	897	658	451	347	**38795**	**31664**	**23227**	**15920**	**12249**
	15:32:38	342	125	101	66	54	**12073**	**4413**	**3565**	**2330**	**1906**
	15:32:44	36	27	23	15	12	**1271**	**953**	**812**	**530**	**424**
	15:32:50	10	9	9	4	3	**353**	**318**	**318**	**141**	**106**
	15:32:56	4	2	1	0	0	**141**	**71**	**35**	**0**	**0**
	15:33:02	0	0	0	0	0	**0**	**0**	**0**	**0**	**0**

Remarks : Recovered.